

1 10A NCAC 14E .0325 is proposed for adoption under emergency procedures as follows:

2

3 **10A NCAC 14E .0325 LABORATORY SERVICES**

4 (a) Each clinic shall have the capability to provide or obtain laboratory tests required in connection with the procedure  
5 to be performed.

6 (b) The governing authority shall establish written policies requiring examination by a pathologist of all surgical  
7 specimens except for those types of specimens that the governing authority has determined do not require examination.

8 (c) Each patient shall have the following performed and a record of the results placed in the patient's medical record  
9 prior to the abortion:

10 (1) pregnancy testing, except when a positive diagnosis of pregnancy has been established by  
11 ultrasound;

12 (2) anemia testing (hemoglobin or hematocrit); and

13 (3) Rh factor testing.

14 (d) Patients requiring the administration of blood shall be transferred to a local hospital having blood bank facilities.

15 (e) The clinic shall maintain a manual in a location accessible by employees, that includes the procedures, instructions,  
16 and manufacturer's instructions for each test procedure performed, including:

17 (1) sources of reagents, standard and calibration procedures, and quality control procedures; and

18 (2) information concerning the basis for the listed "normal" ranges.

19 (f) The clinic shall perform and document, at least quarterly, calibration of equipment and validation of test results.

20

21 History Note: Authority G.S. 143B-10; S.L. 2023-14, s. 2.4;

22 Emergency Adoption Eff.