1	10A NCAC 15	.0110 is proposed for readoption with substantive changes as follows:
2		
3	10A NCAC 15	.0110 PROHIBITED USES
4	(a) Hand held	fluoroscopic screens shall not be used.
5	(b) Shoe fitting fluoroscopic devices shall not be used.	
6	(c) Effective February 1, 1981, plastic pointed position indicating devices on intraoral dental systems shall not be	
7	used.	
8	(d) Effective F	ebruary 1, 1983, mechanical timers on intraoral dental machines shall not be used.
9	(e) Dental fluo:	roscopy without image intensification shall not be used.
10	(f) Non intensi	fied photofluorographic equipment shall not be used.
11	The agency prohibits the use of the following:	
12	<u>(1)</u>	demonstration or training of radiation machines or sources of radiation without providing
13		engineered or administrative protective controls to ensure exposure to radiation does not exceed
14		dose limits in Rule .1601(a) of this Chapter;
15	<u>(2)</u>	hand-held radiation machines used for diagnostic exams, ordered by a licensed practitioner as defined
16		in Rule .0103(7) of this Section in the diagnosing or treatment of human or animal diseases, except
17		for dental hand-held equipment authorized for use by the agency;
18	(3)	hand-held fluoroscopic screens;
19	<u>(4)</u>	shoe-fitting fluoroscopic devices;
20	(5)	dental fluoroscopy without image intensification; and
21	(6)	non-intensified photofluorographic equipment.
22		
23	History Note:	Authority G.S. 104E-7; <u>104E-12(a);</u>
24		Eff. February 1, 1980;
25		Amended Eff. June 1, 1989;
26		Transferred and Recodified from 10 NCAC 3G .2211 Eff. January 4, 1990;
27		Transferred and Recodified from 15A NCAC 11 .0110 Eff. February 1, 2015.2015;
28		Readopted Eff. May 1, 2025.