



NC DEPARTMENT OF
**HEALTH AND
HUMAN SERVICES**

ROY COOPER • Governor
MANDY COHEN, MD, MPH • Secretary
MARK PAYNE • Director, Division of Health Service Regulation

VIA EMAIL ONLY

December 10, 2021

Tommy Williams
tommy.williams@unchealth.unc.edu

No Review

Record #: 3758
Date of Request: December 8, 2021
Facility Name: Johnston Health
FID #: 943290
Business Name: Johnston Memorial Hospital Corporation
Business #: 1052
Project Description: Acquire a CT scanner
County: Johnston

Dear Ms. Culver:

The Healthcare Planning and Certificate of Need Section, Division of Health Service Regulation (Agency) received your correspondence regarding the project described above. Based on the CON law **in effect on the date of this response to your request**, the project as described is not governed by, and therefore, does not currently require a certificate of need. If the CON law is subsequently amended such that the above referenced proposal would require a certificate of need, this determination does not authorize you to proceed to develop the above referenced proposal when the new law becomes effective.

This determination is binding only for the facts represented in your correspondence. If changes are made in the project or in the facts provided in the correspondence referenced above, a new determination as to whether a certificate of need is required would need to be made by this office.

Please do not hesitate to contact this office if you have any questions.

Sincerely,

Kim Meymandi, Project Analyst

Micheala Mitchell, Chief

cc: April Culver, Johnston Health
Acute and Home Care Licensure and Certification Section, DHSR
Construction Section, DHSR
Radiation Protection Section, DHS

NC DEPARTMENT OF HEALTH AND HUMAN SERVICES • DIVISION OF HEALTH SERVICE REGULATION
HEALTHCARE PLANNING AND CERTIFICATE OF NEED SECTION

LOCATION: 809 Ruggles Drive, Edgerton Building, Raleigh, NC 27603
MAILING ADDRESS: 809 Ruggles Drive, 2704 Mail Service Center, Raleigh, NC 27699-2704
<https://info.ncdhhs.gov/dhsr/> • TEL: 919-855-3873

AN EQUAL OPPORTUNITY / AFFIRMATIVE ACTION EMPLOYER



December 3, 2021

VIA ELECTRONIC MAIL

Ms. Micheala Mitchell, Chief
Ms. Kim Meymandi, Project Analyst
NC Department of Health and Human Services
Division of Health Service Regulation
Certificate of Need Section
2704 Mail Service Center
Raleigh, North Carolina 27699-2704

RE: No Review Request by Johnston Health to Acquire a CT Scanner to be located in
Johnston Memorial Hospital

Dear Ms. Mitchell and Ms. Meymandi:

Johnston Health provides this letter as notice to the Department of Health and Human Services, Division of Health Service Regulation, Certificate of Need Section that it plans to acquire a new CT scanner to be located at Johnston Memorial Hospital in Smithfield. Johnston Health seeks confirmation that the purchase of this equipment does not implicate certificate of need ("CON") review and does not require a CON.

Johnston Health does not believe the proposal constitutes a "new institutional health service" as that term is defined in N.C. Gen. Stat. § 131E-176(16). CT scanners are not regulated by Certificate of Need law, and none of the services to be provided via the CT scanner will be a new institutional health service. Further, the total capital expenditure for this acquisition (including the cost to renovate the space where the scanner will be located) will be well under the \$2 million cost threshold for major medical equipment as defined by N.C. Gen. Stat. § 131E-176(16)(14o). A vendor quote for the CT scanner is attached as Exhibit 1, and total projected costs are attached as Exhibit 2.

For these reasons, Johnston Health does not believe that the acquisition of a new CT scanner requires a certificate of need. We respectfully request written confirmation from the Agency that the project described herein does not implicate CON review and does not require a CON.

Please don't hesitate to contact me should you need any additional information or have any questions or concerns.

Sincerely,

A handwritten signature in black ink, appearing to read "Tommy G. Williams", is written over a white background.

Tommy G. Williams, President and CEO



November 17, 2021
 Quote Number: **2006338796.7**
 Customer ID: **1-2310HD**
 Agreement Expiration Date: **15-Feb-22**

Johnston Memorial Hospital
 509 N Brightleaf Blvd
 Smithfield, NC 27577-4407

This Agreement (as defined below) is by and between the Customer and the GE Healthcare business ("GE Healthcare"), each as identified below for the sale and purchase of the Products and/or Services identified in this Quotation, together with any applicable schedules referred to herein ("Quotation"). "Agreement" is this Quotation and either: (i) the Governing Agreement identified below; or (ii) if no Governing Agreement is identified, the GE Healthcare Terms and Conditions and Warranties that apply to the Products and/or Services identified in this Quotation. In the event of conflict, the Quotation supersedes.

GE Healthcare can withdraw this Quotation at any time before Customer: (i) signs and returns this Quotation or (ii) provides evidence of Quotation acceptance satisfactory to GE Healthcare ("Quotation Acceptance"). On Quotation Acceptance, this Agreement is the complete and final agreement of the parties relating to the Products and/or Services identified in this Quotation. There is no reliance on any terms other than those expressly stated or incorporated by reference in this Agreement and, except as permitted in this Agreement, no attempt to modify will be binding unless agreed to in writing by the parties. Modifications may result in additional fees and cannot be made without GE Healthcare's prior written consent.

Handwritten or electronic modifications on this Agreement (except an indication of the form of payment, Customer purchase order number and signatures on the signature blocks below) are void.

Governing Agreement:	GEHC Standard Terms Apply
Terms of Delivery	FOB Destination
Billing Terms	100% billing at Ship Completion (Fulfillment) / Delivery
Payment Terms	Due On Receipt-30 Days
Total Quote Net Selling Price	\$411,740
Sales and Use Tax Exemption	No Certificate on File

IMPORTANT CUSTOMER ACTIONS:

Please select your planned source of funds. Source of funds is assumed to be cash unless you choose another option. Once equipment has been shipped, source of funds changes cannot be allowed.

- Cash
- GE HFS Loan
- GE HFS Lease
- Other Financing Loan
- Other Financing Lease
- Provide Finance Company Name

The parties have caused this Agreement to be executed by their authorized representative as of the last signature date below.

Johnston Memorial Hospital

Signature: _____

Print Name: _____

Title: _____

Date: _____

Purchase Order Number, if applicable

GE Precision Healthcare LLC, a GE Healthcare business

Signature: David Kozloff

Title: Imaging Account Manager

Date: November 17, 2021



November 17, 2021
 Quote Number: **2006338796.7**
 Customer ID: **1-2310HD**
 Agreement Expiration Date: **15-Feb-22**

To Accept This Quotation

Please sign and return this quotation together with your Purchase Order to:

Name: David Kozloff
Email: david.kozloff@ge.com
Phone:
Fax:

Name: Jim Benecki
Email: jim.benecki@ge.com
Phone: (615) 390-3634
Fax: (910) 401-1049

Payment Instructions

Please **remit** payment for invoices associated with this quotation to:

GE Precision Healthcare LLC
P.O. Box 96483
Chicago, IL 60693

FEIN: 83-0849145

Johnston Memorial Hospital

Addresses:

Bill To: JOHNSTON MEMORIAL HOSPITAL

JOHNSTON MEMORIAL HOSPITAL, ACCOUNTS PAYABLE PO BOX 1376
 SMITHFIELD, NC, 27577

Ship To: JOHNSTON MEMORIAL HOSPITAL

509 N BRIGHTLEAF BLVD SMITHFIELD, NC, 27577-4407

To Accept This Quotation

- Please sign the quote and any included attachments (where requested).
- If requested, please indicate your form of payment.
- If you include a purchase order, please make sure it references the following information:
 - The correct Quote number and Version number above
 - The correct Remit To information as indicated in **"Payment Instructions"** above
 - Your correct SHIP TO and BILL TO site name and address
 - The correct Total Price as indicated above

Upon submission of a purchase order in response to this quotation, GE Healthcare requests the following to evidence agreement to contract terms: Signature page on quote filled out with signature and P.O. number **** OR**** Verbiage on the purchase order must state one of the following:

(i) Per the terms of Quotation # _____, (ii) Per the terms of GPO # _____; (iii) Per the terms of MPA# _____; or (iv) Per the terms of SAA # _____.

Include applicable quote/agreement number with the reference on the purchase order. In addition, Source of Funds (choice of Cash/Third Party Load or GE HFS Lease Loan or Third Party Lease through _____), must be indicated, which may be done on the Quote Signature



November 17, 2021
Quote Number: **2006338796.7**
Customer ID: **1-2310HD**
Agreement Expiration Date: **15-Feb-22**

Page (for signed quotes), or the Purchase Order (where quotes are not signed) or via a separate written source of funds statement (if provided by GE Healthcare)."

Catalog Item Details

Line	Qty.	Catalog	
1	1	S7640HH	GoldSeal Revolution EVO EX, with 675 lb table and ASiR

Revolution EVO is the next generation volume CT with a compact design and advanced technologies enabling you to see fine anatomical details, providing a pathway to a quick, confident diagnosis and delivering improved image quality across the entire body. Our innovative iterative reconstruction technologies are designed to reduce noise levels, improve low-contrast detectability and reduce dose for all patients. Additional Smart Dose technologies like organ dose modulation and XR-29 capabilities help you monitor, measure and manage your dose delivery.

Your new system will include the following new h/w and s/w options:

Table slicker
 Footswitch cover/slicker
 Rear Gantry Control Panel
 STRAPS AUTO TRATION
 ARM SUPPORT ASSY
 OC Monitors Eizo
 LONG CABLE COLLECTOR
 BODY STRAP
 CJ CHAIR
 CT DESK
 BIG CABINET
 ENGLISH KYBD KIT GSCB
 GLOBAL SCAN CONTROL BOX
 LOW PROFILE HEAD HOLDER
 CATHETER BAG HOLDER GT
 TABLE TRAY IV POLE
 GT GRANITE base COVER Kit
 Shp Coll-CT660 M40
 REAR CABLE COVER
 VOLUME VIEWER ON OC
 Low Dose CT Lung Screening Option with Indication
 72kW option
 Prospective Exam Split
 ASiR Option kit
 64ch-64sl option
 0.4 sec rotation option
 Workflow Pkg
 UW Protocol - Rev EVO

Revolution EVO Smart Flow technologies are designed to help you improve productivity by streamlining user workflow and access to information, enabling you to perform more studies in less time and manage your patient flow up to 40% more efficiently.

Clarity Imaging Chain

The Clarity imaging chain is a new data acquisition system that is integrated with the photo diode and provides the following benefits:

- Cable free between ASIC and Diode, and has a capability to reduce electric noise
- Up to 90% less heat compared with previous GE technology
- Improved signal to noise up to 44% compared with previous GE technology
- Optimized collimator to reduce scatter dose, noise and artifacts
- Performix40 Plus X-ray tube provides less focus movement

ASiR

ASiR iterative reconstruction technology may enable reduction in pixel noise standard deviation (a measurement of image noise).

The ASiR algorithm may allow for reduced mA in the acquisition of images, thereby reducing the dose required. ASiR also may enable improvement in low contrast detectability. In clinical practice, the use of ASiR may reduce CT patient dose depending on the clinical task, patient size, anatomical location and clinical practice. A consultation with a radiologist and physicist should be made to determine the appropriate dose to obtain diagnostic image quality for the particular clinical task.

Smart Dose

Intelligent technology designed to help you acquire high-quality images using lower doses of radiation, contributing to more accurate diagnoses and lower exposures for patients. Smart Dose includes dose management tools such as:

- Organ Dose Modulation (ODM): ODM provides a reduction of radiation dose via X-ray tube current modulation for sensitive tissues, such as breasts or eyes
- Compliant with the NEMA XR 25, and XR 29 standards
- Adult and Pediatric reference protocols
- Dose Check - Patient pre-scanning monitoring and alerts. Receive notifications and alerts if your predetermined dose levels will be exceeded. Dose check is based on standard XR 25-2010 published by The Association of Electrical and Medical Imaging Equipment Manufacturers (NEMA)
- Dose Reporting: CTDIvol, DLP, Dose Efficiency are displayed to the user during scan prescription and at the end of the exam. The CTDIvol, DLP, and Phantom size used to calculate dose is automatically saved once the user selects End Exam
- DICOM Structured Dose Report generates a CT Dose Report, which can enable tracking of dose (CTDIvol and DLP) for the patient by the hospital radiation tracking system
- 3D mA Modulation utilizing SmartmA and AutomA: 3D mA Modulation allows you to personalize protocols and optimize dose for every patient - large and small
- Dynamic Z-axis tracking: Dynamic Z-axis tracking provides automatic and continuous correction of the x-ray beam shape to block unused x-ray at the beginning and end of a helical scan to reduce unnecessary radiation

DoseWatch Explorer

A web based dose management solution that allows you to track, analyze, identify, and optimize patient dose. This is an introductory dose management software application that provides you secure access, via any PC with internet access, to dose and protocol data from this system. An InSite connection to the system and completion of the registration process is required to use the DoseWatch Explore application. For US and Canadian Customers, this quotation includes access to the DoseWatch Explore application for a period concurrent with the system warranty.

Smart Flow

Designed to help you improve productivity and patient experience by streamlining your workflow and access to information. Smart Flow technologies includes:

- Silent design of Revolution EVO gantry allows significant reduction of audible noise compared with previous GE technology
- Xtream multi-purpose touch display that provides basic patient information, exam information, instructional videos and distraction videos
- Default patient positioning
- One stop scanning mode
- Image Check: Real-time reconstruction, up to 55 images are reconstructed and available per second
- 10 PMR's: Prospectively prescribe up to 10 multiphase reconstructions and easily prioritize which one you need first.
- Direct MPR with Auto-Batch feature, affording automatic real-time direct reconstruction and transfer of fully corrected multi-planar images, also allows users to move from routine 2D review to prospective 3D image review of axial, sagittal, coronal, and oblique planes while enabling automated protocol-driven batch reformats to be created and networked to their desired reading location
- Exam Split
- Volume Viewer on console

Dual Energy: Acquire back to back axial or helical scans of the same anatomy at two different X-ray energies (kVp's). The acquired dual energy data can be post-processed on the console or AW workstation using the Add/Sub function to gain additional clinical information.

IQ Enhance pitch booster - Scan a chest in as fast as two seconds with 175 mm/sec acquisition speed to help shorten patient breath-holds while maintaining image quality. Requires 0.35 second rotation speed capability to achieve 175mm/sec.

System Components:

- Advanced slip ring design
- Aperture: 70 cm
- Maximum SFOV: 50 cm
- Tilt: +/- 30 degrees
- Multi-purpose LCD touch screen display with workflow features
- Integrated start scan button with countdown timer
- Performix40 Plus liquid metal bearing tube
- Heat storage capacity: 7.0 MHU
- Dual Focal Spots:
 - Small Focal Spot: 0.7 (W) x 0.6 (L) Nominal Value; (IEC 60:1993)
 - Large Focal Spot: 0.9 (W) x 0.9 (L) Nominal Value; (IEC 60:1993)
- 72 kW
- kV: 80, 100, 120, 140
- mA: 10 to 560 mA, 5 mA increments
- Full 360° rotational scans: 0.4, 0.5, 0.6, 0.7, 0.8, 0.9, 1.0 second (Optional 0.35 second routine)

Clarity HiLight Detector

- 64 slice system
- 40 mm of coverage @ 0.625 mm slices
- 98% absorption efficiency
- Clarity DAS (Data Acquisition System)

Revolution EVO computer system

Patient table

The 2000 patient table has the following features

- Maximum table load: 675 lbs
- Horizontal speed: 1 – 175 mm/s
- Scannable range: 2,045 mm
- Scout scannable range: 2,000 mm
- Vertical range: 425 – 990 mm
- Elevation speed: 12.3 – 23.2 mm/s

General Electric Company reserves the right to make changes in specifications and features shown herein, or discontinue the product described at any time without notice or obligation.

Laser alignment devices contained within this product are appropriately labeled according to the requirements of the Center for Devices and Radiological Health

GoldSeal shipment notes:

Includes new monitors, new gantry/table leveling pads

Warranty

This product, including the accompanying X-ray tube, comes with a one year warranty.

Availability

Since Preowned Equipment may be offered simultaneously to several customers, its sale to you is subject to availability and subject to prior sale at the time you offer to purchase it. If the equipment is no longer available, (1) GE will attempt to identify other preowned Equipment in our inventory that meets your needs, and (2) if substitute equipment is not acceptable to you, GE will cancel your order and refund any deposit you have paid us for the canceled order.

Asterisk*: Trademark of General Electric Company

Line	Qty.	Catalog	
2	1	B7660EW	SmartScore Package

The SmartScore package provides ECG-gated hardware for both prospective and retrospective gating for coronary artery calcium scoring.

Xtream 12" Gantry and Operator Console ECG Trace: The ECG trace provided by the Ivy monitor will be displayed on the CT gantry and operator console with this option. Allowing the user to display the live trace of the patient's heart rate and display the actual location of the window of time when the image is being acquired. It will provide easy access to patient cardiac output status and assist in providing visual feedback for optimum acquisition start.

The Ivy Monitor comes in this calcium scoring package. It will be used to monitor patient cardiac output and synchronize acquisition with that output.

Calcium scoring analysis software is not provided with this package.

Line	Qty.	Catalog	
3	1	B7880MR	SmartMAR option

SmartMAR (Metal Artifact Reduction) software helps reduce photon starvation, beam hardening and streak artifacts caused by high Z materials in the body, such as hip implants.

The clarity of SmartMAR images is addressing the challenges posed by metal artifacts, helping clinicians accurately contour targets and critical organs.

MAR offers:

Exceptional image quality.

SmartMAR is based on the latest in GE Healthcare smart technology, which uses a novel three-step, sinogram-based iterative algorithm.

Streamlined workflow.

SmartMAR requires only one scan, making the process of obtaining a corrected image fast and efficient.

Dose conscious.

SmartMAR requires only one acquisition.

Patient comfort.

The efficient, single-scan process helps to reduce patient time inside the scanner.

Versatility.

SmartMAR is designed to enhance clarity across a range of images including scans of hip implants, dental fillings, screws and other metal objects.

Line	Qty.	Catalog	
4	1	B7868FM	SmartView Fluoro Option

SmartView software™ provides continuous, real-time CT fluoroscopy with a nominal image lag of only 0.20 seconds and reconstruction at up to 24 fps (3 view ports at 8fps each) with in-room viewing and manual x-ray control. SmartView provides tilted or nontilted imaging for performing biopsies and other interventional procedures with coverage up to 15mm.

The intuitive user interface provides six user-selectable display layouts, in-room image review and WW/WL control.

The image display supports single or multiple real-time images, a free viewport, and timers for remaining and accumulated exposure time. The display control panel provides roam, zoom, magnify, measurement, annotation, grid, image orientation and save screen image review capabilities.

Reconstruction modes, 3i - 24 fps (3 view ports at 8fps each) or 1i - 12fps, may be used to create the following image slice thicknesses with 340 x 340 matrix images for all scan fields of view.

- 1.25mm (16 & 64 channel system only)

- 2.5mm
- 5mm
- 7.5mm
- 10mm

Line	Qty.	Catalog	
5	1	B78552CA	CT Operator Console Desk

The Freedom workspace is an ergonomic working environment specifically designed for use with the GE Healthcare imaging systems. The sleek table design enables the efficient use of space while enhancing clinical workflow and technologist comfort.

The Freedom workspace provides a minimalist footprint to improve patient visibility and giving the user easier access to patients in the imaging suite.

It offers sit/stand and horizontal/vertical monitor flexibility. It can also help reduce noise and heat with remote location options of the console. The non-adjustable Freedom workspace version is 1300mm long x 895mm wide x 850mm height and weighs 55.8kg.

Line	Qty.	Catalog	
6	1	B7660B	Chair

Chair for CT scanner

Line	Qty.	Catalog	
7	1	B76952RE	CT Interventional H/W Kit

The CT intervention kit provides the hardware required for CT interventional procedures. This kit includes the in-room Monitor with suspension arm, Hand Held Controller, X-ray Exposure Foot Pedal and Cradle Handle required for in-room acquisition control and image review. The hand held controller provides the operator with the ability to prepare and perform interventional CT procedures, to turn alignment lights on and off, to move the cradle, review images and adjust the window width/level; and turn x-ray on via the foot switch.

Requires either SmartStep or SmartView to perform CT interventional procedures

Line	Qty.	Catalog	
8	1	B77292CA	CT Service Cabinet

Service cabinet for system accessories storage

Line	Qty.	Catalog	
9	1	R23053AC	Standard Service License

GE Healthcare has reclassified its service tools, diagnostics and documentation into various classes (please refer to the Service Licensing Notification statement at the beginning of this Quotation). The Standard License provides access to service tools used to perform basic level service on the Equipment and is included at no charge for the warranty period.

Line	Qty.	Catalog	
------	------	---------	--

10 1 E4502KZ **Liebert GXT4 10kVA 208Y/120V 2-phase CT partial UPS**

Line	Qty.	Catalog	
11	1	E4502BB	CT Main Disconnect and UPS Control 380-480V 50 60Hz 90A

Main Disconnect Panel (MDP) UL 90A 400/480V 50/60Hz 3 phases for CT, PET and PETCT

The (Main Disconnect and UPS Control Panel serves as the main facility power disconnect source installed ahead of the CT system PDU. On systems where the optional partial system UPS is included in the system, the panel provides NEC mandated UPS emergency power-off control function via a UPS control cable included with the UPS. The optimized design PDB saves time, installation labor, and valuable mounting space by consolidating the main circuit breaker, control power source and required warning lights into a compact factory manufactured panel. The panel provides short circuit protection, overload protection and National Electrical Code and Canadian Electrical Code required emergency shutdown for the system. The 24-volt low voltage controls all power, using either the panel cover mounted EMERGENCY OFF push button or the remote EMERGENCY OFF push button included with each system. The PDB is painted to match the imaging system for a total coordinated system appearance. Available in a combination surface/semi-flush mounted enclosure. The system provides stock availability of otherwise special-order devices, saving time and installation costs.

Benefits

- The System Main Disconnect saves time, installation labor, and valuable mounting space by consolidating the main circuit breaker, the feeder overcurrent devices, magnetic contactors and UPS emergency power-off into one compact panel
- The system provides stock availability of otherwise special-order devices, saving time and installation costs
- Reduces installation time and cost by eliminating delays in obtaining individually enclosed components and by eliminating on site assembly
- UPS emergency power-off functions are included for future, partial system UPS addition.
- Disconnects system power on first loss of incoming power, preventing damage to system components
- Provides a standardized platform for UPS or other future GE engineered modifications or upgrades
- Main power disconnect operating handle can be padlocked in the OFF position for servicing safety and OSHA lock out/tag out
- The door has provisions for padlocking
- Enclosure door is interlocked with ON / OFF disconnect handle to prevent unauthorized access if disconnect is in the ON position

Features

- Optional partial system UPS provides clean uninterrupted power to the system computer, maintaining system integrity during power loss while also providing a solution to power quality problems
- UL, cUL listed, and CE labeled
- Supplied with low voltage, cover mounted Push to Stop, Twist to Restore pushbutton and long-life LED pilot lights
- Provides overcurrent and short circuit protection with GE GuardEON solid-state circuit breakers
- Suitable for use on systems with 25,000A of short circuit current. It is the installer's responsibility to verify that the available short circuit current is 25,000A or less for compliance to all electrical codes
- Emergency-off disconnects power to both the PDU and optional partial system UPS output, per National Electric Code
- Factory wired and tested
- All devices are selected for high reliability and long life
- Panel disconnect provides OSHA lockout / tag out provisions

Remote EPO

- This MDP comes with two normally closed contact blocks attached to the back of the emergency off push button.

Seismic Specifications

- This Panel has been certified by an independent California structural engineer in conformance with the shake testing requirements of ICC-AC 156. The California OSHPD number is OSP-0457-10.
- The seismic performance characteristics are as follows: SDS(g) ≤ 2.56; z/h ≤ 1.0 ; Ip ≤ 1.5

Physical Characteristics

- Dimensions: Height x Width x Depth: 24 x 16 x 7 inches (610 x 407 x 178 mm)
 - Handle depth: 2.75 inches (70 mm)
 - Weight: 46 pounds (21 kg)
- Components supplied with each panel

- The Main Disconnect and UPS Control Panel
- An Installation, Operations & Service Manual
- (2) sets of Emergency Power Off pushbuttons with 2NC on each EPO
- Drawings and Electrical Schematics NOTES:
- Customer is responsible for arranging for installation with a qualified party
- ITEM IS NON-RETURNABLE AND NON-REFUNDABLE

Line	Qty.	Catalog	
12	1	E8016AN	CT Table Slicker with Cushion - 2000 Systems (2-pc Set)

FEATURES/BENEFITS

- Two-piece, sealed slicker cushion set has comfort pads enclosed inside the slicker cover and extender cover
- Durable, clear PVC plastic cover facilitates faster, more thorough cleanup of blood and fluids
- Increase system uptime by protecting table from spills and particulate contaminants
- Thermo-sealed seams and flaps prevent contaminate buildup in hard to clean areas

COMPATIBILITY

- VCT with GT 2000 Table, CT HD750

Line	Qty.	Catalog	
13	1	E8016BA	CT Footswitch Slicker - 2000 & 1700 Systems

The footswitch slicker for CT VCT 2000 and 1700 systems is made of durable, clear PVC plastic that protects the footswitch and facilitates faster, more thorough cleanup of contamination caused by blood and other body fluids. Cover is held securely in place with Velcro.

Line	Qty.	Catalog	
14	1	E80141JA	MEDRAD Stellant FLEX CT Pedestal with Certegra Workstation NO Informatics - includes installation and one year warranty

Dual injector head on pedestal with integral IV pole

- Syringe heat maintainer
- Certegra Workstation with USB drive
- DualFlow software
- ISI-ready software to accept ISI900G integrated injector option†
- Base control unit
- 22.8 m (75 ft) head extension cable
- 7.6m (25 ft) base to display cable
- Power cord
- Product information package
- Operations manuals
- Installation, customer's operational training at time of installation, and one year full on-site warranty in Bayer service countries

Injection Specifications Flow Rate (range & increments):

0.1–10 mL/sec in 0.1 mL Increments

Volume (range & increments):

1 mL to Syringe Capacity in 1 mL Increments
 Programmable Pressure Limit (psi/kPa):
 150 mL and 200 mL Syringe: Choice of 50/345, 100/689, 150/1034, 200/1379, 225/1551, 250/1724, 300/2068, 325/2241
 Scan Reminders:
 0–300 Seconds (5 minutes) in 1 Second Increments
 Pause:
 1–900 Seconds (15 minutes) in 1 Second Increments
 Hold
 Maximum HOLD Time is 20 Minutes
 Syringes (Volume capacity)
 150 mL or 200 mL Sterile Disposable Syringe
 Maximum Number of Phases: 6

Line	Qty.	Catalog	
15	1	W0303CT	TIP CT Scanner 3 Training Program

This training program is designed for customers purchasing a GEHC CT system to include EVO-ES or Discovery RT. GEHC will work with the designated Customer contact to agree upon a reasonable training schedule for a pre-defined group of core technologists that will leverage blended content delivery and may include a combination of onsite days and virtual offerings, to include TiP Virtual Assist, the GEHC Answerline and available on-demand courses ("Virtual Inclusions"). This blended curriculum with multiple delivery platforms promotes learner retention and allows for an efficient and effective skill development.

This program may contain:

- Onsite training (generally 5 days)
- Virtual Inclusions may include:
 - Remote instructor-led training: Instructor leads a remote training session one-on-one or in a group, typically for 1 hour
 - Answerline Support-Access to GEHC experts for clinical, non-emergency applications assistance via phone or by using the iLinq button on the imaging console
 - Tip Virtual Assist-Direct interactive access to a GEHC expert for enhanced support.
 - On Demand courses-On healthcare learning system. Self-paced courses and webinars (CE and non-CE).

Training will be delivered at a mutually agreed upon time between the customer and GE Healthcare (excluding GE Healthcare holidays and weekends), are subject to availability and generally will not exceed 10 days. This training program has a term of twelve (12) months commencing on Acceptance, where all onsite training must be scheduled and completed within twelve (12) months of Acceptance and all Virtual Inclusions also expire at the end of such twelve (12) month period. Additional onsite days may be available for purchase separately.

All GEHC "Training" terms and conditions apply. Given the unique nature of this program, if this program is purchased as part of a purchase under a Governing Agreement, including any Master Purchase Agreement, Group Purchasing Organization Agreement, or Strategic Alliance Agreement, this program shall take precedence over any conflicting training deliverables set forth therein.

Total Quote Subtotal: 411,739.85

Total Quote Net Selling Price: \$411,740

If applicable, for more information on this devices' operating system, please visit GE Healthcare's product security portal at: <https://securityupdate.gehealthcare.com/en/products>

Projected Capital Cost Form

Building Purchase Price	na
Purchase Price of Land	na
Closing Costs	na
Site Preparation	na
Construction/Renovation Contract(s)	\$254,800
Landscaping	na
Architect / Engineering Fees	\$21,000
Medical Equipment	\$411,740
Non-Medical Equipment	
Furniture	
Consultant Fees (specify)	na
Financing Costs	na
Interest during Construction	na
Other (specify)	
Total Capital Cost	\$687,540

CERTIFICATION BY A LICENSED ARCHITECT OR ENGINEER

I certify that, to the best of my knowledge, the projected capital cost for the proposed project is complete and correct.

 Signature of Licensed Architect or Engineer

Date Signed: _____

CERTIFICATION BY AN OFFICER OR AGENT FOR THE PROPONENT

I certify that, to the best of my knowledge, the projected total capital cost for the proposed project is complete and correct and that it is our intent to carry out the proposed project as described.

Stacey Carson
 Signature of Officer/Agent

Date Signed: 11/19/2021

Associate Vice President, Patient Care Services
 Title of Officer/Agent

From: [Meymandi, Kimberly](#)
To: [Waller, Martha K](#)
Subject: FW: [External] Johnston CT Scanner No Review Request
Date: Wednesday, December 8, 2021 3:51:59 PM
Attachments: [Exhibit 1 Vendor Quote.pdf](#)
[CT Scanner No Review Request Dec. 2021.pdf](#)
[Exhibit 2 Cost Estimate.pdf](#)

From: Culver, April <April.Culver@unhealth.unc.edu>
Sent: Wednesday, December 8, 2021 1:22 PM
To: Meymandi, Kimberly <kim.meymandi@dhhs.nc.gov>; Mitchell, Micheala L <Micheala.Mitchell@dhhs.nc.gov>
Subject: [External] Johnston CT Scanner No Review Request

CAUTION: External email. Do not click links or open attachments unless you verify. Send all suspicious email as an attachment to [Report Spam](#).

Ms. Mitchell and Ms. Meymandi:

Please see attached no review request to acquire a CT scanner from Mr. Tommy Williams with attachments. Thank you!

April S. Culver | Vice President, Marketing/Communications and Strategy
Johnston Health
P.O. Box 1376, 509 N. Bright Leaf Blvd., Smithfield, NC 27577
p (919) 938-7198
april.culver@unhealth.unc.edu

----- Confidentiality Notice -----

The information contained in (or attached to) this electronic message may be legally privileged and/or confidential information. If you have received this communication in error, please notify the sender immediately and delete the message.

Email correspondence to and from this address is subject to the North Carolina Public Records Law and may be disclosed to third parties by an authorized State official. Unauthorized disclosure of juvenile, health, legally privileged, or otherwise confidential information, including confidential information relating to an ongoing State procurement effort, is prohibited by law. If you have received this email in error, please notify the sender immediately and delete all records of this email.