

10A NCAC 15 .0808 OTHER EQUIPMENT REQUIREMENTS

(a) RGD's not listed in Rule .0801 of this Section, or that are not able to meet the equipment requirements of either Rule .0806 or .0807 of this Section, shall not be sold, installed, or used prior to the agency completing review of information regarding the RGD and determining if use of the RGD is allowed. The user or manufacturer of the RGD shall submit the following information to the agency for review:

- (1) equipment form for application;
- (2) manufacturer manual;
- (3) description of use;
- (4) operator training;
- (5) a survey in accordance with Rule.0805(d) of this Section;
- (6) an area survey in accordance with Rule.0805(e)(2) of this Section;
- (7) the hazard level associated with use of the RGD; and
- (8) means to achieve radiation protection equivalent to the rules of this Section.

(b) After receiving the information in Paragraph (a) of this Rule, the agency will respond to the applicant in writing within 30 days. Upon review, the agency may require additional information if use of the RGD is allowed.

*History Note: Authority G.S. 104E-7;
Eff. October 1, 2015;
Amended Eff. November 1, 2024.*