

## **10A NCAC 15 .0110      PROHIBITED USES**

The agency prohibits the use of the following:

- (1) demonstration or training of radiation machines or sources of radiation without providing engineered protective barriers or implementing administrative **protective** controls to change work policies, practices, and procedures that will ensure exposure to radiation does not exceed dose limits in Rule .1601(a) of this Chapter;
- (2) hand-held radiation machines used for diagnostic exams, ordered by a licensed practitioner as defined in Rule .0103(b)(7) of this Section, in the diagnosing or treatment of human or animal diseases, except for dental hand-held equipment authorized for use by the agency;
- (3) hand-held fluoroscopic screens;
- (4) shoe-fitting fluoroscopic devices;
- (5) dental fluoroscopy without image intensification; and
- (6) non-intensified photofluorographic equipment.

*History Note:*      *Authority G.S. 104E-7;*  
*Eff. February 1, 1980;*  
*Amended Eff. June 1, 1989;*  
*Transferred and Recodified from 10 NCAC 3G .2211 Eff. January 4, 1990;*  
*Transferred and Recodified from 15A NCAC 11 .0110 Eff. February 1, 2015;*  
*Readopted Eff. May 1, 2025.*