

10A NCAC 15 .0207 is proposed for readoption with substantive changes as follows:

**10A NCAC 15 .0207      ~~ISSUANCE OF NOTICE OF REGISTRATION~~ ADDITIONAL REQUIREMENTS  
TO PROVIDE SERVICES**

~~(a) The agency shall issue a notice of registration upon a determination that an applicant:~~

~~(1) — is qualified by reason of education, training or experience in the use and hazards of radiation sources described in the application for registration;~~

~~(2) — has facilities and equipment which meet the requirements in these Rules;~~

~~(3) — has established a radiation protection program, appropriate to the registered activities, which assures compliance with radiation protection requirements in these Rules; and~~

~~(4) — meets the applicable requirements in this Chapter.~~

~~(b) The agency may, by registration condition or order, when not in conflict with any law, waive any requirement in these Rules or impose requirements with respect to the registrant's receipt, possession, use and transfer of radiation machines as the agency deems appropriate or necessary for compliance with the rules in this Chapter. Such additional requirements are subject to appeal under 15A NCAC 1B .0200.~~

~~(c) The agency may refuse to grant a registration required in Rules .0203 and .0205 of this Section to any applicant who does not possess adequate qualifications or equipment or satisfy the applicable requirements in this Chapter; provided that, before any order is entered denying an application for registration, the agency shall give notice and grant a hearing as provided in G.S. 150B.~~

(a) A person applying for registration to perform Class II or Class IX services for diagnostic radiation output measurements, Class V area radiation surveys and shielding evaluations for diagnostic radiographic and fluoroscopy facilities, or Class VII therapeutic area radiation survey or verification services pursuant to Rule .0205 of this Section shall meet the following additional requirements:

(1) have radiation survey and radiation measurement equipment capable of measuring the radiation energies corresponding to the services requested for authorization;

(2) ensure that the equipment in Subparagraph (a)(1) of this Rule is calibrated annually when a frequency is not recommended by the manufacturer;

(3) submit the following for agency review prior to registration:

(1) a description of the procedures that will be used in performing area radiation surveys including a list of all guides and references to the employed;

(2) a copy of all forms, reports, and documents that will be supplied to registrants;

(3) samples of surveys for each modality requested for registration;

(4) samples of reports of diagnostic radiation output measurements for each modality requested for registration; and

(5) samples of calibration reports for each therapeutic and kV imaging modality requested for registration.

1 (b) A person applying for registration to perform Class VI equipment calibrations shall meet the following  
2 requirements:

- 3 (1) ensure such calibrations are current and traceable to the National Institute of Standards and  
4 Technology;
- 5 (2) license or register radiation sources used for such calibration as required by the rules in this Chapter;
- 6 (3) label the equipment to indicate the date of calibration; and
- 7 (4) maintain records of the calibration.

8 (c) A person applying for registration to perform Class III—shielding designs for diagnostic radiographic facilities,  
9 Class IV shielding designs for diagnostic fluoroscopy facilities, and Class VII therapeutic facilities and shielding  
10 design services shall meet the following additional requirements:

- 11 (1) submit examples of the facility and shielding design which will be provided to registrants;
- 12 (2) submit any technical guides, methodology, occupancy factor rationales, and workload estimation  
13 rationales that will be used; and
- 14 (3) ensure that the facility and shielding design services provided to registrants meet the requirements  
15 in this Chapter.

16  
17 *History Note: Authority G.S. 104E-7;*

18 *Eff. February 1, 1980;*

19 *Amended Eff. June 1, 1993; June 1, 1989;*

20 *Transferred and Recodified from 15A NCAC 11 .0207 Eff. February 1, 2015.*

21 *Readopted Eff. October 1, 2025.*