

1 10A NCAC 15 .1907 is proposed for adoption as follows:

2
3 **10A NCAC 15 .1907 THERAPEUTIC RADIATION MACHINES OF 500 KEV AND ABOVE**

4 a) The licensee shall provide documentation that equipment within this section conforms to the relevant International
5 Electrotechnical Commission standard, documentation of US Food and Drug Administration clearance, or
6 documentation of participation in a research study approved by the licensee's Institutional Review Board.

7 (b) Facility Design Requirements for Therapeutic Radiation Machines Operating above 500 kV. In addition to
8 shielding adequate to meet requirements of Rule .1909 of this Section, the following design requirements are made:

9 (1) Protective Barriers. All protective barriers shall be fixed and permanent with respect to the radiation
10 source and designed to comply with Rules .1601(a)(8) and .1601(a)(15) of this Chapter external to
11 the dedicated space, except for access doors to the treatment space or movable beam interceptors;

12 (2) Control Panel. In addition to other requirements specified within this Section, the control panel shall
13 also:

14 (A) Be located outside the treatment space and complies with Rules .1601(a)(8) and
15 .1601(a)(15) of this Chapter as required; and,

16 (B) Provide an indication of whether radiation is being produced;

17 (3) Include access controls that will prevent unauthorized use of the therapeutic radiation machine;

18 (4) Viewing Systems. Viewing system shall be provided to permit continuous observation of the patient
19 or human research subject following positioning and during irradiation and shall be so located that
20 the operator may observe the patient or human research subject from the treatment control panel.
21 The therapeutic radiation machine shall not be used for patient or human research subject irradiation
22 unless at least one viewing system is operational;

23 (5) Communication Device or Technique. Provision shall be made for continuous two-way
24 communication between the patient or human research subject and the operator at the control panel.
25 The therapeutic radiation machine shall not be used for irradiation of patients or human research
26 subjects unless continuous two-way communication device or technique is possible;

27 (6) Entrances. Treatment space entrances shall be provided with warning lights in a viewable location
28 outside of all entrances, which will indicate when the useful beam is "ON" and when it is "OFF";

29 (7) Entrance Interlocks. Interlocks shall be provided such that all access controls are activated before
30 treatment can be initiated or continued. If the radiation beam is interrupted by any access control, it
31 shall not be possible to restore the machine to operation without activating the access control and
32 reinitiating irradiation by manual action at the control panel;

33 (8) Movable Beam Interceptor Interlocks. If the shielding material in any protective barrier requires the
34 presence of a movable beam interceptor to ensure compliance with Rule .1601(a)(15) of this
35 Chapter, interlocks shall be provided to prevent the production of radiation, unless the beam
36 interceptor is in place, whenever the useful beam is directed at the designated barriers;

1 (9) Emergency Cutoff Switches. At least 1 emergency power cutoff switch shall be located in the
2 radiation therapy room and shall terminate all equipment electrical power including radiation and
3 mechanical motion. All emergency power cutoff switches shall include a manual reset so that the
4 therapeutic radiation machine cannot be restarted from the unit's control console without resetting
5 the emergency cutoff switch; and,

6 (10) Safety Interlocks. All safety interlocks shall be designed so that any defect or component failure in
7 the safety interlock system prevents or terminates operation of the therapeutic radiation machine.

8 (c) Authorized Medical Physicist Support.

9 (1) The services of an Authorized Medical Physicist shall be required in facilities having therapeutic
10 radiation machines. The Authorized Medical Physicist shall be responsible for:

11 (A) Calibrations required by Paragraph (d) of this Rule and radiation safety surveys required
12 by Rule .1904(a) of this Section;

13 (B) Beam data acquisition and configuration for treatment planning, and supervision of its use;

14 (C) Quality assurance, including quality assurance check review required by Paragraph (f) of
15 this Rule.

16 (D) Consultation with the authorized user in treatment planning, as needed; and

17 (E) Perform calculations/assessments regarding medical events.

18 (2) The operating procedures required by Paragraph (d) of this Rule shall also specifically address how
19 the Authorized Medical Physicist is to be contacted for problems or emergencies, as well as the
20 specific actions, if any, to be taken until the Authorized Medical Physicist can be contacted.

21 (d) Operating Procedures.

22 (1) No individual, other than the patient or human research subject, shall be in the treatment space
23 during treatment or during any irradiation for testing or calibration purposes;

24 (2) Therapeutic radiation machines shall not be made available for medical use unless the requirements
25 of Rule .1904(a) of this Section, and Paragraphs (e), (f) and (g) of this Rule have been met;

26 (3) Therapeutic radiation machines, when not in operation, shall be secured to prevent unauthorized use
27 pursuant to Rules .1601(a)(32) and (33) of this Chapter;

28 (4) When a patient or human research subject must be held in position for radiation therapy, mechanical
29 supports or immobilization devices shall be used;

30 (5) A copy of the current operating and emergency procedures shall be maintained at the therapeutic
31 radiation machine control console.

32 (e) Acceptance Testing, Commissioning and Calibration Measurements. Acceptance testing, commissioning, and
33 calibration of a therapeutic radiation machine subject to this Rule shall be performed by, or under the direct supervision
34 of, an Authorized Medical Physicist:

35 (1) Acceptance testing and commissioning shall be performed in accordance with current published
36 recommendations from a recognized national professional association with expertise in the use of
37 therapeutic radiation technologies, that includes the American Association of Physicists in

1 Medicine, the American College of Radiology and the American Society for Radiation Oncology.
2 In the absence of a protocol published by a national professional association, the manufacturer's
3 protocol or equivalent quality, safety, and security protocols, shall be followed.

4 (2) A licensee authorized to use a therapeutic radiation machine for medical use shall perform
5 calibration measurements on each therapeutic radiation machine:

6 (A) Before the first medical use of the unit; and

7 (B) Before medical use under the following conditions: Whenever spot-check measurements
8 indicate that the output, for each specific mode and energy, differs by more than 5 percent
9 from the output obtained at the last calibration, following reinstallation of the therapeutic
10 radiation machine in a new location, following any repair of the therapeutic radiation
11 machine that would likely impact the radiation output beyond the normal range of expected
12 fluctuation; and

13 (C) At intervals not to exceed annually.

14 (3) To satisfy the requirement of Paragraph (d) of this Rule, an authorized medical physicist shall design
15 and implement a calibration procedure for each radiation therapy machine which is consistent with
16 the specifications recommended by the manufacturer of the equipment and consistent with
17 nationally recognizable standards. The calibration procedure shall be designed to ensure accurate
18 patient or human research subject treatments, in accordance with the written directive and treatment
19 plan. The calibration procedure shall include, but not be limited to, the following:

20 (A) Accuracy of output measurements to within $\pm 5\%$ of radiations used medically; and,

21 (B) Evaluation and accuracy of auxiliary systems, such as motion tracking or gating and image
22 guidance, used during patient or human research subject treatments.

23 (f) Independent Verification of Therapeutic Radiation Machine Output

24 (1) In addition to the calibration required by Paragraph (e) of this Rule, the licensee shall have the
25 outputs, for all clinically used radiations, independently verified:

26 (A) Within 90 days of first clinical use of a new installation;

27 (B) Within 90 days of first clinical use following a reinstallation in a new location; and,

28 (C) Biennially, thereafter.

29 (2) Verification may be obtained by:

30 (A) the authorized medical physicist irradiating dosimeters from an AAPM Accredited
31 Dosimetry Calibration Laboratory; or,

32 (B) evaluation by an independent registered qualified expert using an independent dosimetry
33 system meeting Rule .1908 of this Section.

34 (3) A licensee shall maintain a record of each independent verification of therapeutic radiation machine
35 output for three (3) years. The record must include:

36 (A) If obtained by Part (e)(2)(A) of this Rule: The date of the irradiation, the date of the analysis
37 by the dosimetry center, the name, address and contact information for the AAPM

1 Accredited Dosimetry Calibration Laboratory, and the results of the independent
2 verification.

3 (B) If obtained by Part (e)(2)(B) of this Rule: The date of the calibration, The manufacturer's
4 name, model number, and serial number of the therapeutic radiation machine, auxiliary
5 systems, and the instruments used to calibrate the units, the results and an assessment of
6 the independent verification, and the name of the independent registered qualified expert
7 who provided the independent verification.

8 (g) Quality Assurance Checks.

9 (1) Periodic quality assurance checks shall be performed on therapeutic radiation machines subject to
10 this Rule, which are capable of operation at greater than or equal to 500 kV.

11 (2) To satisfy the requirement of Subparagraph (f)(1) of this Rule, quality assurance checks shall meet
12 the following requirements:

13 (A) The licensee shall perform quality assurance checks, to include ensuring the proper
14 function of requirements outlined in Paragraphs (d) and (e) of this Rule, in accordance with
15 written procedures established by the Authorized Medical Physicist; and

16 (B) The quality assurance check procedures shall specify the frequency at which tests or
17 measurements are to be performed. The quality assurance check procedures shall specify
18 that the quality assurance check shall be performed during the calibration specified in
19 Paragraph (d) of this Rule. The acceptable tolerance for each parameter measured in the
20 quality assurance check, when compared to the value for that parameter determined in the
21 calibration specified in Paragraph (d) of this Rule, shall be stated.

22 (3) The cause for a parameter exceeding a tolerance set by the Authorized Medical Physicist shall be
23 investigated and corrected before the system is used for patient or human research subject
24 irradiation;

25 (4) Whenever a quality assurance check indicates a significant change in the operating characteristics
26 of a system, as specified in the Authorized Medical Physicist's quality assurance check procedures,
27 the system shall be recalibrated as required by Paragraph (d) of this Rule;

28 (5) The licensee shall use the dosimetry system described in Rule .1908 of this Chapter to make the
29 quality assurance check required by Paragraph (f) of this Rule;

30 (6) The licensee shall maintain a record of each quality assurance check required by (f) of this rule for
31 three (3) years. The record shall include: the date of the quality assurance check; the manufacturer's
32 name, model number, and serial number of the therapeutic radiation machine; the manufacturer's
33 name; model number and serial number for the instrument(s) used to measure the radiation output
34 of the therapeutic radiation machine; and the signature of the individual who performed the periodic
35 quality assurance check.

36
37 *History Note: Authority G.S. 104E-7;*

