

1 10A NCAC 15 .2004 is adopted as published in 39:19 NCR 1225-1262 as follows:

2
3 **10A NCAC 15 .2004 GENERAL TECHNICAL REQUIREMENTS FOR FACILITIES USING**
4 **VETERINARY THERAPEUTIC RADIATION MACHINES**

5 (a) Protection Surveys:

6 (1) The licensee shall ensure that radiation shielding surveys of all new facilities, and existing facilities
7 not previously surveyed are performed with an operable radiation measurement survey instrument
8 calibrated in accordance with Rule .2008 of this Section. The radiation protection survey shall be
9 performed by, or under the direction of, an Authorized Medical Physicist or a qualified expert, and
10 shall verify that, with the therapeutic radiation machine in a "BEAM-ON" condition:

11 (A) Radiation levels in restricted areas are not likely to cause personnel exposures more than
12 the limits specified in Rule .1601(a)(8) of this Chapter; and

13 (B) Radiation levels in unrestricted areas do not exceed the limits specified in Rule
14 .1601(a)(15) of this Chapter.

15 (2) In addition to the requirements of Subparagraph (a)(1) of this Rule, a radiation protection survey
16 shall also be performed:

17 (A) After making any change in the treatment room shielding;

18 (B) After making any change in the location of the therapeutic radiation machine within the
19 treatment room;

20 (C) After relocating the therapeutic radiation machine;

21 (D) After changes in occupancy of surrounding areas; or

22 (E) Before using the therapeutic radiation machine in a manner that could result in increased
23 radiation levels in areas outside the external beam radiation therapy treatment room.

24 (3) The survey record shall include: the date of the measurements; the reason the survey is required; the
25 manufacturer's name; model number and serial number of the therapeutic radiation machine; the
26 instruments used to measure radiation levels; a plan of the areas surrounding the treatment room
27 that were surveyed; the measured dose rate at several points in each area expressed in microsieverts
28 or millirems per hour; the calculated maximum level of radiation over a period of one week for each
29 restricted and unrestricted area; and the signature of the individual responsible for conducting the
30 survey;

31 (4) If the results of the surveys required by this Paragraph indicate any radiation levels in excess of the
32 limits specified in Parts (1)(A) or (B) of this Paragraph, the licensee shall disable the machine from
33 use, label clearly, and not use the unit:

34 (A) Except as may be necessary to repair, replace, or test the therapeutic radiation machine, the
35 therapeutic radiation machine shielding, or the treatment room shielding; or

36 (B) Until the licensee has received a specific exemption from the Agency.

1 (b) Modification of Radiation Therapy Unit or Room Before Beginning a Treatment Program. If the survey required
2 by Paragraph (a) of this rule indicates that an individual in an unrestricted area may be exposed to levels of radiation
3 greater than those permitted by Rule .1601 of this Chapter, before beginning the treatment program the licensee shall:

4 (1) Either equip the unit with beam direction interlocks or add additional radiation shielding to ensure
5 compliance with Rule .1601 of this Chapter;

6 (2) Perform the survey required by Paragraph (a) of this Rule again; and

7 (3) Include in the report required by Paragraph (d) of this Rule the results of the initial survey, a
8 description of the modification made to comply with Subparagraph (b)(1) of this Rule, and the
9 results of the second survey; or

10 (4) Receive an amended license issued by the agency that authorizes radiation levels in unrestricted
11 areas greater than those permitted by Rule .1601 of this Chapter.

12 (c) Radiation Measuring Equipment. The licensee shall have, when required, appropriate and operable radiation
13 measuring equipment available for use and calibrated in accordance with Rule .2008 of this Section. Radiation
14 measuring equipment includes, but is not limited to, dosimetry systems, survey instruments, and other radiation
15 measuring devices used in planning, guiding, and administering radiation.

16 (d) Reports of External Beam Radiation Therapy Surveys and Measurements. The licensee for any therapeutic
17 radiation machine subject to Rules within this subpart shall furnish a copy of the records required in Paragraphs (a)
18 and (b) of this Rule to the Agency within 30 days following completion of the action that initiated the record
19 requirement.

20
21 History Note: Authority G.S. 104E-7;

22 Eff. October 1, 2025.